



M A X T E R
GLOVE MANUFACTURING SDN BHD
(229862-H)

LOT 6070

Jalan Haji Abdul Manan, 6th Miles Off Jalan Meru
41050 Klang, Selangor, Malaysia
Tel: 603-33929888 (8 lines) Fax: 603-33923328
E-MAIL: info@maxter.com.my

Date: 31st May 2021

To Whom It May Concern

EU DECLARATION OF CONFORMITY

We, **MAXTER GLOVE MANUFACTURING SDN. BHD.** located at **Lot 6070, Jalan Haji Abdul Manan, 6th Miles off Jalan Meru, 41050 Klang**, declares under our sole responsibility that the medical devices described hereafter as :-

- **Non Sterile Powder Free Latex Examination Gloves, Basic UDI-DI: 955 500211 6378**
 - **Non Sterile Powdered Latex Examination Gloves, Basic UDI-DI: 955 500211 6361**
 - **Non Sterile Powder Free Nitrile Examination Gloves, Basic UDI-DI: 955 500211 6385**
-
- Are in conformity with the general safety and performance requirements of Annex I Medical Device Regulation (EU) 2017/745 for Class I medical devices.
 - Classification: Class I based on Rule 5 transient use, Annex VIII of Medical Device Regulation (EU) 2017/745.
 - Are in conformity with the national standard transposing harmonized standard EN 455.
 - The gloves are manufactured according to ISO 9001:2015 and EN ISO 13485:2016 Quality Management Systems and certified by Notified Body, SGS United Kingdom Ltd Systems & Services Certification.
 - Our Authorized Representative is Supermax Healthcare (Europe) Limited, 38 Main Street, Swords Co. Dublin, Ireland K67 E0A2.

Klang, Selangor
Malaysia



Yap Peak Geeh
QA & Regulatory Affairs Manager

**MAXTER GLOVE MANUFACTURING SDN BHD
GLOVE SPECIFICATION**

1. Product : (5.5g) Smooth Surface Powdered Latex Examination Gloves
 2. Material : Natural Rubber Latex
 3. Type : Non Sterile
 4. Glove Design & Feature : Ambidextrous, Smooth Surface
 5. Cuff : Beaded
 6. Usage : For Single Use Only
 7. Colour : Beige
 8. Packaging : 100 ± 2 pcs/inner, 10 or 20 inners/carton

9. Dimensions (Available in 5 sizes) :

Size	Glove Weight (g)	Palm Width (mm)		Length (mm)	
		ASTM D3578	MAXTER	ASTM D3578	MAXTER
XS	4.6 ± 0.2	70 ± 10	70 – 79	Min. 220	Min. 240
S	5.0 ± 0.2	80 ± 10	80 – 89	Min. 220	Min. 240
M	5.4 ± 0.2	95 ± 10	90 – 99	Min. 230	Min. 240
L	5.8 ± 0.2	110 ± 10	100 – 109	Min. 230	Min. 240
XL	6.3 ± 0.2	120 ± 10	110 – 119	Min. 230	Min. 240

10. Thickness :

Location of Thickness Measurement	Single Wall (mm)	
	ASTM D3578	MAXTER
Finger	Min 0.08	Min 0.10
Palm	Min 0.08	Min 0.09
Cuff	N/A	Min 0.08

11. Physical Properties :


	Before Aging		After Aging	
	ASTM D3578	MAXTER	ASTM D3578	MAXTER
Tensile Strength (MPa)	Min. 18	Min. 18	Min. 14	Min. 14
Elongation (%)	Min. 650	Min. 650	Min. 500	Min. 500
Force at Break (N)	N/A	Min. 6	N/A	Min. 6


12. Protein Content : Max. 200 µg/dm²
 13. Powder Level : Max. 10 mg/ dm²
 14. Compliance Standards : ASTM D3578, EN455 Part 1, 2, 3 &4.

15. Inspection Criteria (Sampling & Inspection ISO 2859) :

	Inspection Level	AQL
Watertight Test	G1	1.5
Major Defects	G1	2.5
Minor Defects	G1	4.0
Dimensions	S2	4.0
Physical Properties	S2	4.0

16. Quality Assurance : ISO 9001 certified by SGS (UK) and SIRIM (Malaysia)
 ISO 13485 certified by SGS (UK)

Prepared by : 
 Yap Peak Geeh
 QA & RA Manager

Approved by : 
 Dato' Seri Stanley Thai
 Group MD

